

DC-30

Color Doppler Ultrasound System

Specification

Release V01.00.00



1 System Overview

1.1 Application

- Abdomen
- Obstetrics
- Gynecology
- Small parts
- Vascular
- Urology
- Musculoskeletal
- Orthopedics
- Emergency Medicine
- Nerve
- Cardiology
- Pediatrics
- Others

1.2 Transducer types

- Curved array
- Linear array
- Phased array transducer
- 4D Volume transducer

1.3 Imaging modes

- B-Mode
- Tissue Harmonic and PSH (Phase Shift Harmonic Imaging)
- M-Mode
- Color Doppler Imaging
- Power Doppler Imaging/Directional PDI
- Pulsed Wave Doppler
- Continuous Wave Doppler
- Free Xros M™ (Anatomical M-mode)
- TDI
- UWN(Ultra Wideband Non-linear) Contrast Imaging
- Natural Touch Elastography Imaging
- Smart 3D™ (Freehand 3D)
- Real-time 4D
- iScape™ View (Panoramic Imaging)

1.4 Standard features

- B-Mode
- THI and PSH™ (Phase Shift Harmonic Imaging)
- M-Mode

- Color Doppler Imaging
- Power Doppler Imaging and Directional PDI
- Pulsed Wave Doppler
- iClear™ (Speckle Suppression Imaging)
- iBeam™ (Spatial Compounding Imaging)
- iTouch™ (Auto Optimization)
- Zoom/iZoom™ (Full Screen Zoom)
- FCI (Frequency Compounding Imaging)
- B steer
- ExFOV Imaging
- iStation™
- iVision™
- Integrated 1TB hard drive
- 3 active transducer connectors
- DVD-RW Driver
- 3 USB ports
- Auto Doppler Calculation
- Shared Service Package
- Medsight
- iStorage (Direct Network Storage)
- Tutorial function: iScanHelper
- Smart Installment Reminder

1.5 Optional features

- iScape™ View
- Auto IMT (Auto Intima-Media Thickness Evaluation)
- Continuous Wave Doppler
- Free Xros M(Anatomical M-mode)
- TDI
- UWN(Ultra Wideband Non-linear) Contrast Imaging
- Natural Touch Elastography Imaging
- Smart 3D
- Real-time 4D
- iPage™ (Multi-Slice Imaging)
- Nerve Application Package
- Emergency & Critical Application Package
- DICOM
- ECG function

- Independent Gel warmer
- Built-in battery

1.6 Language support

- Software display: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Turkish, Norwegian, Serbian, Swedish, Hungarian
- Keyboard input: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Icelandic, Norwegian, Swedish, Finnish, Turkish, Danish
- Control panel overlay: Chinese, Italian, Portuguese, Spanish, German, Russian, French, Czech, Polish
- User manual: English, Chinese

2 Physical Specification

2.1 Dimension and weight

- Width: 520mm
- Height:
 - 1265~1415mm (with adjustable height)
 - 1315mm (without adjustable height)
- Depth: 670mm
- Weight: approx. 55kg (with battery)

2.2 Monitor

- 15-inch/17-inch high resolution color LED monitor
- Resolution: 1024×768 (15-inch); 1280×1024 (17-inch)
- Digital on-screen display of brightness and contrast controls
- Viewing angle: 140° (15-inch) up/down and left/right; 178°(17-inch) up/down and left/right
- Tile/Rotate independent adjustment:
 - Tilt angle range:110 degrees
 - Rotate angle range: 180 degrees

2.3 Transducer port and holder

- 3 active transducer ports

- Removable transducer holder

2.4 Electrical power

- AC adapter Input:
 - Voltage:100-240V~
 - Frequency: 50/60Hz
 - Power consumption: 420VA
- Built-in Battery: Lithium-ion Battery 14.8V, 6600mAh

2.5 Operating Environment

- Ambient temperature: 0-40 °C
- Relative humidity: 30%-85% (no condensation)
- Atmospheric pressure: 700hPa-1060hPa

2.6 Storage & Transportation Environment

- Ambient temperature: -20~55 °C
- Relative humidity: 20%-95% (no condensation)
- Atmospheric pressure: 700hPa-1060hPa

3 User Interface

3.1 Control panel

- Power/Battery indicator
- Alphanumeric keys
- Function keys
- Knobs
- Backlit keys, ensuring accurate work in the dark room
- 8-segment TGC control
- 4 programmable keys, available for user-defined functions
- Trackball, sensitivity adjustment
- Key brightness adjustment
- Up/down of control panel
 - Down/Up range: 150mm

3.2 System boot-up

- Boot-up in about 46 sec
- Shut down in about 19sec
- Boot-up from standby mode in about 23 sec
- Enter standby mode in about 13 sec

3.3 Comments

- Support text input and arrow

- Adjustable text size, arrow size and direction
- Support home position
- Covers various application
- User customizable

3.4 Bodymark

- More than 140 bodymarks for versatile application
- User customizable

3.5 Screen information*

- Common info:
 - Mindray logo
 - Hospital name
 - Exam date
 - Exam time
 - Acoustic power
 - Mechanical index
 - Tissue thermal index
 - ID, Last name, First Name, Middle initial, Gender, Age
 - Transducer model
 - Operator
 - TGC Curve
 - Focus position
 - Thumbnail
 - Imaging parameters
 - iScanHelper guidance

*Not all items are listed in this part, detail info please refer to user manual

4 Imaging Parameters

4.1 Overview

- Echo-enriched Beamformer
- Up to 27,648 channels
- 4-beam forming

4.2 B-mode

- Display formats: Single(B), Dual(B+B), Quad(4B)
- iClear™
- iBeam™
- iTouch™
- Frequency
- B steer: available on linear transducers

- ExFOV
- Depth
- Acoustic output power
- TGC
- Dynamic range
- Gain
- Focus number
- Focus position
- FOV (Field of View)
- Line density
- Persistence
- Horizontal Scale
- L/R flip
- U/D flip
- TSI (Tissue Specific Imaging)
- Gray Map
- Tint map
- Gray Invert
- Auto merge
- Middle Line
- LGC

4.3 THI and PSH

- Available on all types of transducers
- Patent PSH technology, obtains purer harmonic, better contrast resolution
- iClear™ available

4.4 M-mode

- Display formats
- Acoustic output power
- Dynamic range
- Gain
- Speed
- M soften
- Tint map
- Gray Map
- Edge enhance

4.5 Color Doppler Imaging

- Dual live
- iTouch™
- Frequency
- Steer
- Acoustic output power
- Gain

- ROI size/position
- Scale
- Baseline
- Wall filter
- PRF
- Packet size
- Smooth
- B/C Align
- Priority
- Map
- Invert
- Persistence
- Line density

4.6 Power Doppler Imaging

- Dual live
- Support directional PDI
- Frequency
- Acoustic output power
- Dynamic range
- Gain
- ROI size/position
- Steer
- Scale
- Wall filter
- PRF
- Packet size
- Smooth
- B/C Align
- Priority
- Map
- Persistence
- Line density

4.7 PW Mode/ CW Mode

- Display formats
- iTouch™
- Frequency
- Acoustic power
- PW velocity
- CW velocity
- Sample volume size
- Sample gate depth
- Scale
- Baseline
- PW Steer

- Audio
- PW PRF
- CW PRF
- Gain
- Dynamic range
- Speed
- Wall filter
- Invert
- Angle
- Quick angle
- Gray map
- Tint map
- Time/frequency resolution
- Auto calc
- Trace area
- HPRF
- Duplex/Triplex
- Auto calc cycle
- Auto calc param

4.8 Smart 3D (option)

- Smart 3D™
 - Angle
 - Distance
 - Display formats
 - Reset
 - Quick Rotation
 - Render type
 - Accept VOI
 - Render
 - Direct
 - Threshold
 - Opacity
 - Smooth
 - Bright
 - Contrast
 - Tint
 - Current window
 - iClear
- Edit
 - Rotation control
 - Tool
 - Other operations

4.9 Real-time 4D

- MPR/VR

- Threshold
 - Smooth
 - iClear
 - Current window
 - Display formats
 - Brightness
 - Contrast
 - Opacity
 - Reset
 - Quick Rotation
 - Render type
 - Accept VOI
 - Render
 - Direct
 - Tint
 - Slice
 - Edit
 - Rotation control: X, Y, Z axis
 - Tool: inside contour, outside contour; inside rect, outside rect
 - Format: Single, Dual, Quad, A4:1
 - Other operations: undo, undo all
 - Auto Rotation
- 4.10 iPage
- Slices number
 - Spacing
 - Slice Position
 - Line Direction
 - Ref. Plane
 - Display format
 - Adjust Slice
 - Range Position
 - Ref. Image
 - Reset Orientation
- 4.11 iScape™ View (option)
- Panoramic imaging
 - Available on all transducers
 - Imaging length
 - Tint map
 - Rotation
- 4.12 Tissue Velocity/Energy Imaging (included in TDI option)
- Available on phased array transducer
- Dual live: side by side displays B and B+TVI
 - Max velocity
 - Max. frame rate
 - PRF
 - Acoustic output power
 - Gain
 - Dynamic range
 - ROI size/position
 - Scale
 - Baseline
 - Wall filter
 - Packet size
 - Smooth
 - B/C Align
 - Priority
 - Map
 - Invert
 - Persistence
 - Line density
- 4.13 Tissue Velocity Doppler (included in TDI option)
- Available on phased array transducer
 - Display formats
 - Duplex/Triplex
 - Max velocity
 - Sample volume size
 - Sample gate depth
 - Scale
 - Baseline
 - Audio
 - PRF
 - Gain
 - Dynamic range
 - Speed
 - Wall filter
 - Invert
 - Angle correction
 - Quick angle
 - Gray map
 - Tint map
 - Time/frequency resolution
- 4.14 Tissue Velocity Motion (included in

- TDI option)
 - Display formats
 - Acoustic output power
 - Dynamic range
 - Gain
 - Speed
 - M soften
 - Gray Map
 - Edge enhance
- 4.15 Free Xros M™ (option)
 - Display formats
 - Support 1 line
 - Sweep speeds
 - M Tint map
 - Gray Map
- 4.16 Natural Touch Elastography
 - Available on 75L38P in Thyroid, Breast or MSK mode
 - Stress compensation technology reduces deeper tissue artifacts, obtains more uniform stress throughout whole field
 - Stress indicator
 - Display format
 - Elasto Map
 - Smooth
 - Invert
 - Opacity
- 4.17 UWN Contrast Imaging™* (option)
 - Ultra Wideband Non-linear (UWN) contrast imaging technology, which provides exceptional contrast agent detecting capability, not only extracts second harmonic, but also non-linear fundamental signals
 - Available on 35C50P in Adult ABD mode
 - Timer1
 - Timer2
 - Pro capture: captures prospective image less than 480s
 - Retro capture: captures retrospective image less than 120s
 - Dual live
- Destruct
- iClear
- Mix
- Mix map
- Persistence
- Dynamic range
- Gray map
- Tint map
- Supports U/D Flip and L/R Flip
- Rotation
- HlmgPos
- Line density
- Destruct AP
- Destruct time
- 4.18 Zoom
 - iZoom™
 - Full screen zoom
 - Normal image, Zoom standard area, Zoom image area, 3 steps
 - Spot zoom (write zoom) 0.8-10x
 - Pan zoom (read zoom) 0.8-10x
- 4.19 iBeam™
 - Spatial compound imaging
 - 3 angles maximum
 - Available on all convex and linear transducers
- 4.20 iClear™
 - Speckle suppression imaging
 - Available for B, 3D, 4D
- 4.21 iTouch™
 - Auto image optimization
 - B-mode
 - Color
 - Power
 - PW
- 4.22 B steer
 - Only for linear transducers
- 4.23 ExFov
 - Extended field of view
 - Available for all convex, linear and volume transducers
- 4.24 QSave
 - Quick save image parameter setting after image adjustment done

- Support Save, Save as, Restore
- 4.25 iScanHelper
- Tutorial function as a guidance to show basic scanning skill with graphic of probe position, schematic of anatomy and example clinical image
 - Support ABD, OB/GYN, Thyroid, Breast and Testicle applications

5 Cine Review and Post Processing

5.1 Cine review

- Available in all modes
- Frame by frame manual cineloop review or auto playback with variable speed
- Independent cine review in 2D Dual and Quad mode one by one
- Maximum cine memory is up to 12,394 frames and PW mode up to 181s
- Cine length: 1-60s
- Frame compare: compare different frames for one cine in dual format
- Cine compare: compare two or more than two cines in dual or quad format
- Skip to first and skip to last: one keystroke review the first or last frame
- Start point and end point: selectable

5.2 Post Processing

- B-mode:
 - TGC
 - Gain
 - Zoom
 - Gray map
 - Tint map
 - Flip
 - iClear
 - H Scale
 - LGC
 - Gray invert
 - Auto merge

- M-mode:
 - Gray map
 - Tint map
- Color/Power:
 - Invert
 - Baseline
 - Map
 - Priority
 - Smooth
 - Dual Live
 - Dynamic Range (Power)
- PW:
 - Gain
 - Baseline
 - Audio
 - Angle
 - Dynamic Range
 - Angle correction
 - Quick angle
 - Invert
 - Gray map
 - Tint map
 - Auto calc
 - Auto Calc Cycle
 - Auto Calc Param
 - Trace Area

6 Measurement/Analysis and Report*

6.1 Generic measurements

- 2D-mode
 - Depth
 - Distance
 - Angle
 - Area: Ellipse, Trace, Spline, Cross
 - Trace Length
 - Double Distance
 - Parallel
 - Volume: 3-Distance, Ellipse, Ellipse + Distance)
 - Length Ratio
 - Area Ratio
 - IMT

- B Histogram
 - B Profile
 - Volume Flow
 - Color Velocity
 - M-mode
 - Distance
 - Time
 - Slope
 - Heart Rate
 - Velocity
 - Doppler mode
 - D Velocity
 - Time
 - Heart Rate
 - Acceleration
 - D Trace
 - PS/ED
 - Volume Flow
 - Automatic Doppler Spectrum Analysis
 - Heart cycle pre-settable (1, 2, 3, 4, 5)
 - Automatic tracing in real-time
 - User configurable display of items
 - Support PI, RI, TAMAX, TAMEAN, Volume Flow calculations
 - Appropriate factory setting according to applications
- 6.2 Clinical option measurement package
- Abdominal
 - Liver
 - Renal Length
 - Renal Height
 - Renal Width
 - Renal Cortical Thickness
 - Adrenal Length
 - Adrenal Height
 - Adrenal Width
 - Common bile duct
 - Portal Vein Diameter
 - Common hepatic duct
 - Gallbladder Length
 - Gallbladder Height
 - Gallbladder wall thickness
 - Pancreatic duct
 - Pancreatic head
 - Pancreatic body
 - Pancreatic tail
 - Spleen
 - Aorta Diameter
 - Iliac Diameter
 - Pre-void Bladder Length
 - Pre-void Bladder Height
 - Pre-void Bladder Width
 - Post-void Bladder Length
 - Post-void Bladder Height
 - Post-void Bladder Width
 - Ureter
 - Renal Artery Origin
 - Arcuate Artery
 - Segmental Artery
 - Interlobar Artery
 - Renal Artery
 - Main Renal Artery
 - Renal Vein
 - Aorta
 - Celiac Axis
 - Superior Mesenteric Artery
 - Common Hepatic Artery
 - Hepatic Artery
 - Splenic Artery
 - Inferior Vena Cava
 - Portal Vein
 - Main Portal Vein
 - Hepatic Vein
 - Left Hepatic Vein
 - Right Hepatic Vein
 - Middle Hepatic Vein
 - Splenic Vein
 - Superior Mesenteric Vein
 - Renal Volume
 - Pre-void Bladder Volume
 - Post-void Bladder Volume
 - Micturated Volume
 - Gynecology
 - Uterine Length
 - Uterine Height
 - Uterine Width
 - Uterine Cervix Length

- Uterine Cervix Height
- Uterine Cervix Width
- Endometrium Thickness
- Ovary Length
- Ovary Height
- Ovary Width
- Follicle 1~16 Length
- Follicle 1~16 Width
- Follicle 1~16 Height
- Follicle Volume
- Ovary Volume
- UT Volume
- Uterus Body
- UT-L/ CX-L
 - Obstetrics
- EGestational Sac Diameter
- Yolk Sac
- Crown Rump Length
- Nuchal Translucency
- Biparietal Diameter
- Occipital Frontal Diameter
- Head Circumference
- Abdominal Circumference
- Femur Length
- Abdominal Transversal Diameter
- Anteroposterior Abdominal Diameter
- Cerebellum Diameter
- Cist Magna
- Lateral Ventricle Width
- Hemisphere Width
- Outer Orbital Diameter
- Inter Orbital Diameter
- Humerus Length
- Ulna Length
- Radius Length
- Tibia Length
- Fibula Length
- Clavicle Length
- Length of Vertebrae
- Middle Phalanx Length
- Foot Length
- Ear Length
- Anteroposterior trunk diameter
- Transverse trunk diameter
- Fetal Trunk Cross-sectional Area
- Thoracic Diameter
- Heart Circumference
- Thoracic circumference
- Umbilical Vein Diameter
- Fetal kidney Length
- Matrix Kidney Length
- Cervical Length
- Amniotic Fluid
- Nuchal Fold
- Orbit
- Placental Thickness
- Gestational Sac Diameter 1
- Gestational Sac Diameter 2
- Gestational Sac Diameter 3
- Amniotic Fluid 1
- Amniotic Fluid 2
- Amniotic Fluid 3
- Amniotic Fluid 4
- Left Ventricular Internal Diameter at End-diastole
- Left Ventricular Internal Diameter at End-systole
- Left Ventricular Diameter
- Left Atrium Diameter
- Right Ventricular Internal Diameter at End-diastole
- Right Ventricular Internal Diameter at End-systole
- Right Ventricular Diameter
- Right Atrium Diameter
- Interventricular Septal Thickness at End-diastole
- Interventricular Septal Thickness at End-systole
- Interventricular Septal Thickness
- Aorta Diameter
- Main Pulmonary Artery Diameter
- Right Ventricular Outflow Tract Diameter
- Right Ventricular Outflow Tract Diameter
- Left Ventricular Area
- Left Atrium Area

- Right Ventricular Area
- Right Atrium Area
- Heart area
- Facial Angle
- Mitral Valve diameter
- Pulmonary valve Diameter
- Ascending Aorta Diameter
- Descending Aorta Diameter
- Ductus Arteriosus Diameter
- Tricuspid valve Diameter
- Left pulmonary Artery Diameter
- Right pulmonary Artery Diameter
- Inferior vena cava Diameter
- Aorta Valve Diameter
- Mean Gestational Sac Diameter
- AFI
- Estimated Fetal Weight
- Estimated Fetal Weight 2
- HC/AC
- FL/AC
- FL/BPD
- AXT
- CI
- FL/HC
- HC(c)
- HrtC/TC
- TCD/AC
- LVW/HW
- LVD/RVD
- LAD/RAD
- AoD/MPAD
- LAD/AoD
- Fetal Heart Rate
- Left ventricular short-axis diameter at end diastole
- Left ventricular short-axis diameter at end systole
- Right ventricular short-axis diameter at end diastole
- Right ventricular short-axis diameter at end systole
- Interventricular septal thickness at en diastole
- Interventricular septal thickness at en systole
- Umbilical Artery
- Ductus Venosus
- Placenta Artery
- Middle Cerebral Artery
- Fetal Aorta
- Descending Aorta
- Uterine Artery
- Ovarian Artery
- Fetal Heart Rate
 - Cardiology
- Left Atrium Diameter
- Left Atrium major Diameter
- Left Atrium minor Diameter
- Right Atrium major Diameter
- Right Atrium minor Diameter
- Left Ventricular major Diameter
- Left Ventricular minor Diameter
- Right Ventricular major Diameter
- Right Ventricular minor Diameter
- Left Atrium area
- Right Atrium area
- Left Ventricular area at end-diastole
- Left Ventricular area at end-systole
- Right Ventricular area at end-diastole
- Right Ventricular area at end-systole
- Left Ventricular Internal Diameter at end-diastole
- Left Ventricular Internal Diameter at end-systole
- Right Ventricular Diameter at end-diastole
- Right Ventricular Diameter at end-systole
- Left Ventricular Posterior wall thickness at end-diastole
- Left Ventricular Posterior wall thickness at end-systole
- Right Ventricular Anterior wall thickness at end-diastole
- Right Ventricular Anterior wall thickness at end-systole
- Interventricular Septal thickness at end-diastole

- Interventricular Septal thickness at end-systole
- Aorta Diameter
- Aorta arch Diameter
- Ascending Aorta Diameter
- Descending Aorta Diameter
- Aorta Isthmus Diameter
- Aorta ST junct Diameter
- Aorta Sinus Diameter
- Ductus Arteriosus Diameter
- Previous ductal Diameter
- Posterior ductal Diameter
- Aortic Valve Cusp Separation
- Left Ventricular Outflow Tract Diameter
- Aorta Valve Diameter
- Aortic Valve Area
- Pulmonary valve Diameter
- Left pulmonary Artery Diameter
- Right pulmonary Artery Diameter
- Main pulmonary Artery Diameter
- Right Ventricular Outflow Tract Diameter
- Mitral Valve diameter
- Mitral Valve area
- Mitral Valve Cusp Separation
- Distance between point E and Interventricular Septum when mitral valve is fully open
- Tricuspid valve Diameter
- Tricuspid Valve Area
- Inferior vena cava inspiration Diameter
- Inferior vena cava expiration Diameter
- Superior vena cava inspiration Diameter
- Superior vena cava expiration Diameter
- Left Coronary Artery
- Right Coronary Artery
- Ventricular Septal defect Diameter
- Atrial Septal defect Diameter
- Patent ductus Arteriosus Diameter
- Patent Oval Foramen Diameter
- Pericardial Effusion at diastole
- Pericardial Effusion at systole
- Heart Rate
- End-diastolic Left Ventricular Measurement
- End-systolic Left Ventricular Measurement
- Left Atrium Diameter/Aorta Diameter
- Aorta Diameter/Left Atrium Diameter
- Left Atrium Diameter
- Left Ventricular Internal Diameter at end-diastole
- Left Ventricular Internal Diameter at end-systole
- Right Ventricular Diameter at end-diastole
- Right Ventricular Diameter at end-systole
- Left Ventricular Posterior wall thickness at end-diastole
- Left Ventricular Posterior wall thickness at end-systole
- Right Ventricular Anterior wall thickness at end-diastole
- Right Ventricular Anterior wall thickness at end-systole
- Interventricular Septal thickness at end-diastole
- Interventricular Septal thickness at end-systole
- Aorta Diameter
- Aorta arch Diameter
- Ascending Aorta Diameter
- Descending Aorta Diameter
- Aorta Isthmus Diameter
- Aorta ST junct Diameter
- Aorta Sinus Diameter
- Left Ventricular outflow tract Diameter
- Aortic valve Cusp Separation
- Left pulmonary Artery Diameter
- Right pulmonary Artery Diameter
- Main pulmonary Artery Diameter
- Right Ventricular outflow tract Diameter
- Amplitude of the Mitral Valve E wave
- Amplitude of the Mitral Valve A wave

- Mitral Valve E-F slope
- Mitral Valve D-E slope
- Amplitude of the Mitral Valve DE wave
- Mitral Valve Cusp Separation
- Distance between point E and the interventricular septum
- Pericardial effusion at diastole
- Pericardial effusion at systole
- Left Ventricular pre-ejection period
- Left Ventricular ejection time
- Right Ventricular pre-ejection period
- Right Ventricular ejection time
- Heart Rate
- End-diastolic Left Ventricular Measurement
- End-systolic Left Ventricular Measurement
- Left Atrium diameter/Aorta diameter
- Aorta Diameter/Left Atrium Diameter
- Mitral Valve Maximum Velocity
- Mitral Valve E-wave Velocity
- Mitral Valve A-wave Velocity
- Mitral Valve E-wave Velocity-Time Integral
- Mitral Valve A-wave Velocity-Time Integral
- Mitral Valve Velocity-Time Integral
- Mitral Valve Acceleration Time
- Mitral Valve Deceleration Time
- Isovelocity Relaxation Time
- Isovelocity Compression Time
- Mitral Valve E-wave Duration
- Mitral Valve A-wave Duration
- Left Ventricular Outflow Tract Velocity
- Left Ventricular Outflow Tract Velocity-Time Integral
- Left Ventricular Outflow Tract Acceleration Time
- Ascending Aorta Maximum Velocity
- Descending Aorta Maximum Velocity
- Aorta Valve Maximum Velocity
- Aorta Valve Velocity-Time Integral
- Left Ventricular Pre-ejection Period
- Left Ventricular Ejection Time
- Aorta Valve Acceleration Time
- Aorta Valve Deceleration Time
- Right Ventricular Ejection Time
- Right Ventricular Pre-ejection Period
- Tricuspid Valve Maximum Velocity
- Tricuspid Valve E-wave Flow Velocity
- Tricuspid Valve A-wave Flow Velocity
- Tricuspid Valve Velocity-Time Integral
- Tricuspid Valve Acceleration Time
- Tricuspid Valve Deceleration Time
- Tricuspid Valve A-wave Duration
- Right Ventricular Outflow Tract Maximum Velocity
- Right Ventricular Outflow Tract Velocity-Time Integral
- Pulmonary Valve Maximum Velocity
- Pulmonary Valve Velocity-Time Integral
- Pulmonary Valve Acceleration Time
- Main Pulmonary Artery Maximum Velocity
- Right Pulmonary Artery Maximum Velocity
- Left Pulmonary Artery Maximum Velocity
- Pulmonary Vein S-wave Flow Velocity
- Pulmonary Vein D-wave Flow Velocity
- Pulmonary Vein A-wave Flow Velocity
- Pulmonary Vein A-wave Duration
- Pulmonary Vein S-wave Velocity-time Integral
- Pulmonary Vein D-wave Velocity-time Integral
- Pulmonary Vein Deceleration Time
- Inferior Vena Cava Inspiration Maximum Velocity
- Inferior Vena Cava Expiration Maximum Velocity
- Superior Vena Cava Inspiration Maximum Velocity
- Superior Vena Cava Expiration Maximum Velocity
- Mitral Valve Regurgitation Maximum Velocity
- Mitral Valve Regurgitation

- Velocity-Time Integral
- Mitral Valve Stenosis Maximum Velocity
- Rate of Pressure change
- Aortic Valve Regurgitation Maximum Velocity
- Aortic Valve Regurgitation Velocity-Time Integral
- Aortic Valve Regurgitation Deceleration Time
- Aortic Valve Regurgitation Pressure Half Time
- Aortic Valve Regurgitation Velocity at end-Diastole
- Tricuspid Valve Regurgitation Maximum Velocity
- Tricuspid Valve Regurgitation Velocity-Time Integral
- Pulmonary Valve Regurgitation Maximum Velocity
- Pulmonary Valve Regurgitation Velocity-Time Integral
- Pulmonary Valve Regurgitation Pressure Half Time
- Pulmonary Valve Regurgitation Velocity at end-Diastole
- Ventricular Septal Defect Maximum Velocity
- Atrial Septal Defect Maximum Velocity
- Patent Ductus Arteriosus Velocity at End-diastole
- Patent Ductus Arteriosus Velocity at End-systole
- Coarctation of Pre-Ductus
- Coarctation of Post-Ductus
- Heart Rate
- Right Atrium Pressure
- Mitral Valve E-Vel/A-Vel
- Mitral Valve Orifice Area (PHT)
- Tricuspid Valve E-Vel/A-Vel
- Tricuspid Valve Orifice Area (PHT)
 - Urology
- Renal Length
- Renal Height
- Renal Width
- Renal Cortical Thickness
- Adrenal Length
- Adrenal Height
- Adrenal Width
- Prostate Length
- Prostate Height
- Prostate Width
- Seminal Vesicle Length
- Seminal Vesicle Height
- Seminal Vesicle Width
- Testicular Length
- Testicular Height
- Testicular Width
- Ureter
- Pre-void Bladder Length
- Pre-void Bladder Height
- Pre-void Bladder Width
- Post-void Bladder Length
- Post-void Bladder Height
- Post-void Bladder Width
- Prostate Mass1 Distance 1~3
- Prostate Mass2 Distance 1~3
- Prostate Mass3 Distance 1~3
- Testis Mass1 Distance 1~3
- Testis Mass2 Distance 1~3
- Testis Mass3 Distance 1~3
- Renal Volume
- Prostate Volume
- Testicular Volume
- Pre-void Bladder Volume
- Post-void Bladder Volume
- Micturated Volume
 - Vascular
- Common Carotid Artery IMT
- Bulbillate IMT
- Internal Carotid Artery IMT
- External Carotid Artery IMT
- Stenosis Diameter
- Stenosis Area
- Intima-Media Thickness
- Common Carotid Artery
- Bulbillate
- Internal Carotid Artery

- External Carotid Artery
- Vertebral Artery
- Innominate Artery
- Subclavian Artery
- Axillary Artery
- Brachial Artery
- Ulnar Artery
- Radial Artery
- Subclavian Artery
- Axillary Vein
- Cephalic Vein
- Basilic Vein
- Ulnar Vein
- Radial Vein
- Common Iliac Artery
- External Iliac Artery
- Common Femoral Artery
- Superficial Femoral Artery
- Popliteal Artery
- Tibial Peroneal Trunk Artery
- Peroneal Artery
- Posterior Tibial Artery
- Anterior Tibial Artery
- Dorsalis Pedis Artery
- Common Iliac Vein
- External Iliac Vein
- Femoral Vein
- Great Saphenous Vein
- Popliteal Vein
- Tibial Peroneal Trunk Vein
- Sural Vein
- Soleal Vein
- Peroneal Vein
- Posterior Tibial Vein
- Anterior Tibial Vein
- Anterior Cerebral Artery
- Middle Cerebral Artery
- Posterior Cerebral Artery
- Ant.communicating br.
- Post.communicating br.
- Basilar Artery
- Internal Iliac Artery
- Deep Femoral Artery
- Basilar Vein

- Brachial Vein
- Internal Iliac Vein
- Common Femoral Vein
- Superficial Femoral Vein
- Deep Femoral Vein
- Small Saphenous Vein
- Ankle Systolic Pressure
- Brachial Systolic Pressure
 - Small Parts
- Thyroid Length
- Thyroid Height
- Thyroid Width
- Isthmus height
- Testicular Length
- Testicular Height
- Testicular Width
- Breast Mass1 d1-d3
- Breast Mass2 d1-d3
- Breast Mass3 d1-d3
- Thyroid Mass1 d1~d3
- Thyroid Mass2 d1~d3
- Thyroid Mass3 d1~d3
- Thyroid Volume
- Superior Thyroid Artery
- Inferior Thyroid Artery
 - Orthopedics
- Hip
- d/D

6.3 IMT

- Intima-Media Thickness measurement
- Automatic detection of IMT when ROI is set
- Support CCA, ICA, ECA, Bulb IMT
- Near wall and far wall detection
- Angle selectable
- Support IMT growth curve

6.4 Report

- Specific report template to the application
- Editable value in report
- Images are selectable
- Titles are pre-settable in setup
- Export as PDF/RTF format

* Not all measurements are listed in this part; For more detailed information please refer to User Manual

7 Exam Storage and Management

7.1 Exam storage

- 1TB hard drive. About 726GB internal hard drive reserved for patient data storage
- Capable of storage up to approximately 248,149 single frames (FRM format)
- Storage area
- Pre-settable: image area, standard area, full-screen
- Image area: 640*480
- Standard area: 800*600
- Full-screen: 1024*768

7.2 Exam management

- iStation™ workstation dedicated for patient exam management
- Patient exam query/retrieve
- Support review of current and past exam
- New exam, Activate exam, Continue exam functions, End exam are available
- Support measurements and calculations on archived exam and images
- Export image as BMP/JPG/TIFF/DCM/FRM format (FRM: system format)
- Export cine as DCM/AVI/CIN format (CIN: system format)
- Support backup/send to USB devices, CD-RW/DVD-RW media

8 Connectivity

8.1 Ethernet Network Connection

- Wired connection

8.2 DICOM 3.0

- DICOM Basic (option)
- Task management

- Print
- Storage
- Storage Commitment
- Media Exchange
 - DICOM Worklist (option)
 - DICOM Modality Performed Procedure Step - MPPS (option)
 - DICOM OB/GYN structure report (option)
 - DICOM Cardiac structure report (option)
 - DICOM Vascular structure report (option)
 - DICOM Query/Retrieve (option)

8.3 iStorage(included in UltraAssist tool)

- Direct network storage tool between ultrasound system and personal computer

8.4 MedSight

- An interactive app that lets you transfer clinical images straight from Mindray Ultrasound system to a smart device, such as mobile phone or tablet PC.
- Needs to be installed on mobile terminal
- Transfer images or clips from system to mobile terminal through WiFi
- Support both IOS and Android powered system.
- For IOS powered smart device: DICOM is mandatory, IOS 5.0 or above; For Android powered smart device: DICOM not necessary

9 Transducers

9.1 Curve array

- 35C50P
 - Application: Abdomen, Gynecology, Obstetrics, Pediatric, Vascular, Urology
 - Bandwidth: 1.8-6.0 MHz(-20dB)
 - Convex Radius: 50mm
 - Biopsy Guide: available, multi angle,

- reusable
- 3C5P
 - Application: Abdomen, Gynecology, Obstetrics, Pediatric, Abdominal, Vascular, Urology
 - Bandwidth: 1.7-6.0 MHz (-20dB)
 - Convex Radius: 50mm
 - Biopsy Guide: available, multi angle, reusable
- 6C2P
 - Application: Small part, Pediatric, Musculoskeletal, Vascular, Nerve
 - Bandwidth: 3.3-11.3MHz(-20dB)
 - Convex Radius: 15mm
 - Biopsy Guide: available, multi angle, reusable
- 6CV1P
 - Application: Gynecology, Obstetrics, Urology
 - Bandwidth: 3.5-12.3MHz(-20dB)
 - Convex Radius: 10mm
 - Biopsy Guide: available, reusable

9.2 Volume curved array

- D6-2P
 - Application: Abdomen, Gynecology, Obstetrics
 - Bandwidth: 1.7-5.6 MHz (-20dB)
 - Convex Radius: 40mm
 - Biopsy Guide: not available

9.3 Linear array

- 75L38P
 - Application: Small part, Musculoskeletal, Vascular, Nerve, Pediatric
 - Bandwidth: 3.4-12.9 MHz (-20dB)
 - Field of View (max): 38 mm
 - Biopsy Guide: available, multi angle, reusable
- 7L4P
 - Application: Small part, Musculoskeletal, Vascular, Nerve, Pediatric
 - Bandwidth: 3.0-13.0 MHz (-20dB)
 - Field of View (max): 38 mm

- Biopsy Guide: available, multi angle, reusable

- 7L5P
 - Application: Small part, Musculoskeletal, Vascular, Nerve, Pediatric
 - Bandwidth: 3.2-11.8 MHz (-20dB)
 - Field of View (max): 52 mm
 - Biopsy Guide: available, multi angle, reusable

9.4 Phased array

- 2P2P
 - Application: Adult Cardiac, Transcranial, Adult Abdomen
 - Bandwidth: 1.5-5.0 MHz (-20dB)
 - Field of View (max): 90°
 - Biopsy Guide: available, multi angle, reusable

10 Peripheral Devices and

Accessories (Option)

- 10.1 Analog Black/white video printer
 - MITSUBISHI P93W-Z
 - SONY UP-X898MD
- 10.2 Analog Color video printer
 - SONY UP-20
 - MITSUBISHI CP910E
- 10.3 Digital Black and White Video Printer
 - SONY D897
- 10.4 Digital Black and White Video Printer
 - SONY UP-D23/25MD
- 10.5 Graph/text printer
 - HP Deskjet Ink Advantage 2020hc
- 10.6 Footswitch
 - USB port: FS-81-SP-2 (1-pedal)
 - USB port: 971-SWNOM (2-pedal)
 - USB port: 971-SWNOM (3-pedal)
 - Support User-definable functions (Freeze, Save, Print)
- 10.7 ECG
 - 6-pin, AHA/IEC, for 3-lead wires
 - ECG wave display: on/off
 - Gain: 0-30

- Sweep speed: 1-6, 1/step
- 10.8 Built-in Battery
- Replaceable and rechargeable lithium battery
 - Continuous work time: about 1.5h in B mode
 - Full battery lasts about 24h in standby mode
 - Empty battery recharged to full in less than 3h
- 10.9 Built-in DVD R/W
- DVD R/W drive

11 System Inputs and Outputs

- 11.1 Video/Audio output
- Video out: 1 port
 - Audio out: 2 ports
 - S-Video out: 1 port
 - VGA out: 1 port
- 11.2 Physio input
- Support ECG signal
 - ECG: 1 port
- 11.3 Other input/output
- USB: 3 ports
 - Ethernet: 1 port
 - Remote control: 1 port

12 Safety and Conformance

- 12.1 Quality standards
- ISO 9001:2008
 - ISO 13485:2003

12.2 Design standards

- EN 60601-1 and IEC 60601-1
- EN 60601-1-2 and IEC 60601-1-2
- EN 60601-2-37 and IEC60601-2-37
- EN ISO 14971 and ISO 14971
- EN ISO10993-1 and ISO10993-1
- EN 62366 and IEC 62366
- EN ISO 17664
- EN 62304 and IEC 62304
- EN 1041
- EN ISO 15223-1

12.3 CE declaration

DC-30 system is fully in conformance with the Council Directive 93/42/EEC Concerning Medical Devices, as amended by 2007/47/EC.. The number adjacent to the CE marking (0123) is the code of the EU-notified body that certified meeting the requirements of Annex II of the Directive.

NOTICE:

Not all features or specifications described in this document may be available in all transducers and/or modes.

Mindray reserves the right to make changes in specifications and features shown herein, or discontinue the product at any time without notice or obligation. Contact Mindray Representative for the most current information